


Dept.	Bar & Rod Br.	No.	20150232	Detect Date	2015.2.14
Sample Condition	M	Heat No.	304E-141219	Size	φ 98.5mm
Grade	Ti6Al4VELI	Detect Project	Microstructural Rating	Analyze Equipment	Leica5000 Metalloscope



TITANEX® 136 Spec.

- ASTM F 136, Ti6Al4V
- ISO 5832-3
- ETTC2 Ed. 2 A1...A7
- Ra max. 1.65
- 290 HV +5/-10 HV
- T/S max 910 MPa

- dia. 98.5/94 (+/-0.1) mm
- THK (+/-0.1) mm

- REACH
- RoHS
- Conflict free Material

Vergleich zu Mustern vom freien Markt:

